

Speckle Interferometry at IMBL - First Results

Monday, 2 December 2019 18:36 (1 minute)

At IMBL inline phase-contrast imaging is typically used for imaging and tomography. This utilises Paganin's phase-retrieval algorithm (Paganin et al. 2002) for boosting the signal to noise of the resultant images. However, this single-image algorithm does not permit the independent extraction of absorption and phase shift arising from a sample, nor can the dark-field signal be obtained.

A number of alternative imaging methods such as grating-based phase-contrast, edge-illumination and speckle interferometry enable the independent extraction absorption, phase and dark-field signals (Mayo & Endrizzi 2018). Grating-based methods and edge-illumination require careful alignment and relative movement of two sets of gratings, or of a grating and detector, with images acquired at a succession of positions to obtain high quality data. Speckle interferometry and related tracking methods are experimentally less demanding, requiring only a mask and the ability to move it repeatably to different positions.

The speckle mask creates a fine scale random pattern in the imaging plane and images are acquired with and without the sample in place. The local differences in intensity, position and contrast of the mask image with and without the sample are used to extract the different signals. For high-resolution speckle interferometry multiple image-pairs are acquired with the mask in different positions (Zdora et al. 2018). However, useful results at lower resolution can be achieved even with a single image pair (Berujon et al 2012).

Here we present preliminary results from our first attempts to implement speckle interferometry at IMBL, including testing of single-image and multi-image methods including tomography.

References

Paganin D, Mayo SC, Gureyev TE, Miller PR, Wilkins SW (2002) Simultaneous phase and amplitude extraction from a single defocused image of a homogeneous object. *Journal of Microscopy-Oxford* 206:33-40

Mayo S., Endrizzi M. (2018) X-Ray Phase Contrast Methods. In: Ida N., Meyendorf N. (eds) *Handbook of Advanced Non-Destructive Evaluation*. Springer, Cham

Zdora MC, Thibault P, Deyhle H, Vila-Comamala J, Rau C, Zanette I (2018) Tunable x-ray speckle-based phase-contrast and dark-field imaging using the unified modulated pattern analysis approach. *J Instrum* 13

Berujon S, Wang H, Sawhney K (2012) X-ray multimodal imaging using a random-phase object. *Physical Review A* 86:063813

Level of Expertise

Experienced Researcher

Do you wish to take part in the poster slam

No

Travel Funding

No

Speakers Gender

Female

Primary author: Dr MAYO, Sheridan (CSIRO Manufacturing,)

Co-authors: BOWDEN, Josh (CSIRO IMT); LI, Thomas (University of Canterbury); Dr PAGANIN, David (Monash); PAVLOV, Konstantin M. (University of Canterbury); MAKSIMENKO, Anton (Australian Synchrotron); KENNEDY, Ben (University of Canterbury); KITCHEN, Marcus (Monash University); STEVENSON, Andrew (Australian Synchrotron/ CSIRO)

Presenter: Dr MAYO, Sheridan (CSIRO Manufacturing,)

Session Classification: Welcome Function