



Contribution ID: 38

Type: Poster

A Feasibility Study of X-Ray Phase-Contrast Mammographic Tomography at the IMBL of the Australian Synchrotron

Friday, 27 November 2015 13:30 (45 minutes)

We present the results of our recent experiment at the Imaging and Medical beamline of the Australian synchrotron intended to contribute to the development and implementation of low-dose, high-sensitivity 3D mammographic phase-contrast imaging, initially at synchrotrons and subsequently in hospitals and medical imaging clinics. We have tested, quantified, evaluated and optimised the effect of such imaging parameters as X-ray energy, source size, detector resolution, sample-to-detector distance, scanning and processing strategies in the case of propagation-based phase-contrast CT, using a plastic phantom simulating relevant breast tissue characteristics. The data are collected using a Hamamatsu CMOS Flat Panel Sensor, utilised in partial scan mode, with the pixel size $100\mu\text{m} \times 100\mu\text{m}$. Analysis of the data revealed the presence of propagation-based phase contrast and demonstrated significant improvement of the quality of phase-contrast CT imaging, compared to conventional (absorption-based) CT, at medically acceptable radiation doses.

Keywords

Tomography phase-contrast synchrotron mammography

Primary author: Dr NESTERETS, Yakov (CSIRO)

Co-authors: Dr STEVENSON, Andrew (CSIRO); Dr JEREMY, Brown (Monash University); Dr LOCKIE, Darren (Maroondah BreastScreen); Mr THOMPSON, Darren (CSIRO); Dr BRUN, Francesco (Elettra-Sincrotrone); Dr TROMBA, Giuliana (Elettra-Sincrotrone); Dr PAVLOV, Konstantin (University of New England); KITCHEN, Marcus (Monash University); Dr MAYO, Sheridan (CSIRO); Dr GUREYEV, Timur (ARC Centre of Excellence in Advanced Molecular Imaging)

Presenter: Mr THOMPSON, Darren (CSIRO)

Session Classification: Poster Session 2

Track Classification: Imaging