



Contribution ID: 58

Type: Oral

## A Comparison of Different Approaches to Image Quality Assessment in Phase-Contrast Mammography

Thursday, 25 November 2021 11:15 (15 minutes)

Propagation-based phase-contrast computed tomography (PB-CT) has the potential to improve breast cancer detection and characterisation compared to established mammography techniques. The aim of this work is to find a quantitative image quality metric which could accurately predict the subjective clinical image quality assessment of PB-CT images made by radiologists as described in Taba et al. [1].

The experimental data analysed in this study included PB-CT scans, which were obtained for 12 full intact mastectomy samples at Imaging and Medical beamline (IMBL) of the Australian Synchrotron at different monochromatic X-ray energies and clinically relevant radiation doses. Quantitative image quality metrics including visibility, signal to noise ratio (SNR), and spatial resolution, were calculated for all PB-CT and conventional CT image sets using the open-source 3D Slicer (<https://www.slicer.org/>) software. For each metric, an objective image quality “score” was generated to match the subjective scoring provided by the radiologists. Weighting factors were then applied to the scores and a weighted contrast to noise to resolution (CNR/res) score was calculated.

The unscaled contrast and spatial resolution scores were both found to have a significant correlation with the radiologists’ scores with R values of 0.9223 and 0.8360 respectively, while SNR had an insignificant correlation, with an R value of -0.6785. The weighted CNR/res score showed a significant correlation to the radiologists’ scores with an R value of 0.9681.

[1] S. T. Taba et al., Academic radiology 28.1 (2021): e20-e26.

Presenter Gender

Man

Do you wish to take part in the Student Poster Slam

Condition of submission

Yes

Students Only - Are you interested in AINSE student funding

Yes

Level of Expertise

Student

## Pronouns

He/Him

Which facility did you use for your research

Australian Synchrotron

Primary authors: REYNOLDS, Jesse (University of Canterbury); BRENNAN, Patrick; GUREYEV, Timur (the University of Melbourne); PAVLOV, Konstantin

Presenter: REYNOLDS, Jesse (University of Canterbury)

Session Classification: Biomedicine, Life science & Food Science

Track Classification: Biomedicine, Life science & Food Science