



Contribution ID: 15

Type: Poster

Data processing technique for the Taipan Be-filter spectrometer

Thursday, 25 November 2021 17:54 (1 minute)

Taipan, the highest flux thermal neutron scattering instrument at ACNS, was originally built as a traditional triple-axis spectrometer. In 2016 a beryllium filter analyser spectrometer was added for increased versatility. The Be-filter acts like a low-energy band-pass filter ideal for investigating lattice dynamics and molecular vibrations over a wide energy range. It is particularly well suited to measuring the motion within materials containing light elements such as hydrogen.

We have successfully created a robust method of treating data from the Taipan filter-analyser and present the method within this poster [1]. The data-treatment process includes correction for the non-linear energy variation of a particular monochromator, removal of higher-order wavelength contamination, and estimation of low-energy multiple-scattering. The steps described here can be utilized by all users of the Australian Nuclear Science and Technology Organisation “Be-filter”—past, present, and future.

[1] G.N. Iles, K.C. Rule, V.K. Peterson, A.P.J. Stampfl and M.M. Elcombe, Rev. Sci. Instrum. 92, 073304 (2021); doi: 10.1063/5.0054786

Presenter Gender

Woman

Do you wish to take part in the Student Poster Slam

Condition of submission

Yes

Students Only - Are you interested in AINSE student funding

Level of Expertise

Expert

Pronouns

Which facility did you use for your research

Primary authors: Dr ILES, Gail (RMIT); RULE, Kirrily (ANSTO); PETERSON, Vanessa (ANSTO); STAMPFL, Anton (Australian Nuclear Science and Technology Organisation); Dr EL-COMBE, Margaret (ANSTO)

Presenter: RULE, Kirrily (ANSTO)

Session Classification: Poster Session

Track Classification: Instruments & Techniques